



Requête alert system

1. Requête exits the printer via sheet of paper.
2. Emitter/detector sensor is triggered, sending signal to ctrl box.
3. Ctrl box turns on Green LED and turns on buzzer cct which sounds buzzer for 1 sec every 2 minutes.
4. Xray tech pushes reset button to clear alert and reset system, proceeds to the xray room to do the requette.